



EZ-SCREEN® CUP Package Insert-II

The EZ-SCREEN® Cup products are rapid qualitative screening assays for the detection of any combination of the following drugs or their metabolites in human urine: Amphetamine, Barbiturates, Benzodiazepines, Cocaine, Methamphetamine, Methadone, Opiates, Oxycodone, Phencyclidine, Propoxyphene, and TH (Cannabinoids). **Configurations of EZ-SCREEN® Cup products can consist of any combination of the tests listed in this insert. Refer to product labeling for the drugs assayed by the kit configuration.**

The adulterant strip is a rapid qualitative screening assay for the detection of Oxidants, Nitrites, and the determination of Specific Gravity and pH values in human urine. It is used to evaluate specimens for adulteration and dilution prior to Drugs of Abuse urine (DAU) testing. The adulterant strip is only for forensic/toxicology use and not for *in vitro* diagnostic applications.

1. INTENDED USE

The EZ-SCREEN® Cup Drugs of Abuse Test is a one-step immunochromatographic test for the rapid, qualitative detection of one or more of the following: Amphetamine, Barbiturates, Benzodiazepines, Cocaine, Methamphetamine, Methadone, Opiates, Oxycodone, Phencyclidine, Propoxyphene, and TH (Cannabinoids) in human urine. EZ-SCREEN® Cup is not for over-the-counter sale. The test detects drug classes at the following cutoff concentrations:

Test ID	Drug Class (calibrator)	Cutoff	Test ID	Drug Class (calibrator)	Cutoff
AM	Amphetamine (d-Amphetamine)	1000 ng/mL	OP	Opiates (Morphine)	2000 ng/mL
BA	Barbiturates (Butalbital)	200 ng/mL	OX	Oxycodone (Oxycodone)	100 ng/mL
BZ	Benzodiazepines (Nordiazepam)	300 ng/mL	PC	Phencyclidine (Phencyclidine)	25 ng/mL
CO	Cocaine (Benzoylecgonine)	300 ng/mL	PP	Propoxyphene (Norpropoxyphene)	300 ng/mL
mA	Methamphetamine (d-Methamphetamine)	1000 ng/mL	TH	Cannabinoids (11-nor-9-carboxy- Δ^9 -THC)	50 ng/mL
MT	Methadone (Methadone)	300 ng/mL			

THE EZ-SCREEN® CUP DRUGS OF ABUSE TEST PROVIDES ONLY A PRELIMINARY ANALYTICAL TEST RESULT. A MORE SPECIFIC ALTERNATE CHEMICAL METHOD MUST BE USED IN ORDER TO OBTAIN A CONFIRMED ANALYTICAL RESULT. GAS CHROMATOGRAPHY / MASS SPECTROMETRY (GC/MS), HIGH PERFORMANCE LIQUID CHROMATOGRAPHY (HPLC) OR LIQUID CHROMATOGRAPHY / TANDEM MASS SPECTROMETRY (LC/MS/MS) ARE THE PREFERRED CONFIRMATORY METHODS. CLINICAL CONSIDERATION AND PROFESSIONAL JUDGMENT SHOULD BE APPLIED TO ANY DRUG OF ABUSE TEST RESULT, PARTICULARLY WHEN PRESUMPTIVE POSITIVE RESULTS ARE OBTAINED.

2. SUMMARY AND EXPLANATION OF THE TEST

Qualitative EZ-SCREEN® Cup Drugs of Abuse Test screens utilize a one-step, solid-phase immunoassay technology to provide a very rapid test requiring no instrumentation. This test may be used to screen urine samples for one or more of the following drug classes prior to confirmatory testing:

The "Amphetamines" are a group of drugs that are central nervous system stimulants. This group includes 'amphetamine' and 'methamphetamine', and related designer drugs like '3,4 Methylendioxyamphetamine', (better known as Ecstasy or MDMA, a psychoactive drug with hallucinogenic effects).

The drug 'Amphetamine' (d-amphetamine) is detected on the device only at the (AM) position. Both the designer drug Ecstasy (mA) 'Methylendioxyamphetamine' and methamphetamine (d-methamphetamine) are detected on the device at the (mA) position. The (mA) antibody does not differentiate between methamphetamine and ecstasy.

Barbiturates (BA) are a group of structurally related prescription drugs that are used to reduce restlessness and emotional tension, induce sleep and to treat certain convulsive disorders.

Benzodiazepines (BZ), a group of structurally related central nervous system depressants, are primarily used to reduce anxiety and induce sleep.

Cocaine (CO) is a central nervous system stimulant. Its primary metabolite is benzoylecgonine.

Methadone (MT) is a synthetic opioid used clinically as a maintenance drug for opiate abusers and for pain management.

Opiates (OP) are a class of natural and semi-synthetic sedative narcotic drugs that include morphine, codeine and heroin.

Oxycodone (OX) (Oxycontin®, Percodan, Percocet) is a semi synthetic narcotic analgesic that is prescribed for moderately severe pain. It is available in both standard and sustained release oral formulations. Oxycodone is metabolized to Oxymorphone and Noroxycodone.

Phencyclidine (PC) is a hallucinogenic drug.

Propoxyphene (PP) is a narcotic analgesic. Its primary metabolite is norpropoxyphene.

Marijuana (TH) is a hallucinogenic drug derived from the hemp plant. Marijuana contains a number of active ingredients collectively known as Cannabinoids.

Many factors influence the length of time required for drugs to be metabolized and excreted in the urine. A variety of factors influence the time period during which drug metabolites are detected in urine; the rate of urine production, the volume of fluid consumption, the amount of drug taken, the urine pH, and the length of time over which drug was consumed. Drinking large volumes of liquid or using diuretics to increase urine volume will lower the drug concentration in the urine and may decrease the detection period. Although the detection period for these drugs varies widely depending upon the compound taken, dose and route of administration and individual rates of metabolism, some general times have been established and are listed below.¹⁻⁵

<u>Drug</u>	<u>Detection Period</u>
Amphetamine Acid Conditions Alkaline Condition	1-3 days 3-10 days
Barbiturates Short-Acting Long-Acting	Up to 6 days Up to 16 days
Benzodiazepines	1-12 days
Cocaine metabolite	Up to 5 days 1 to 3 days typical
Methadone	1-3 days
Methamphetamine Acid Conditions Alkaline Conditions	1-3 days 3-10 days

<u>Drug</u>	<u>Detection Period</u>
Opiates Heroin Morphine Codeine	1 day 1-3 days 1-3 days
Oxycodone	1-3 days
Phencyclidine Single Use Chronic Use	1-8 days Up to 4 weeks
Propoxyphene	Up to 1 week
TH (Cannabinoids) Single Use Chronic Use	1-7 days Less than 30 days typical

The adulterant strip has impregnated reagent test pads that detect specific analytes in human urine. The analytes detected are Oxidants and Nitrites. The strip also approximates the pH and specific gravity values. Urine samples with 'abnormal' values should be submitted to a reference laboratory for additional testing.

Oxidants The detection is based on the oxidative activity of compounds (e.g. chromate salts and/or Bleach) that catalyze the oxidation of an indicator by an organic hydroperoxide producing a blue/orange color. The color intensity is directly proportional to the concentration of Oxidants present in the sample and is observed visually and compared to the color comparator chart to obtain a result.

Nitrites The test is based on the principles of the Griess reaction for the detection of Nitrites. The test pad contains an amine and a coupling component. A red/orange colored azo compound is obtained by diazotization and subsequent coupling. The color intensity is directly proportional to the concentration of Nitrites present in the sample and is observed visually and compared to the color comparator chart to obtain a result.

pH The test pad contains indicators that change colors between pH 2 and pH 11. The color scale gives an approximate indication for pH values between those levels.

Specific Gravity The test pad reacts with ions in urine to indicate concentrations from 1.000 to 1.020. The color changes range from dark green with low ionic concentrations through green to yellow/orange in urines with high ionic concentrations. The color is observed visually and compared to the color comparator chart to obtain an approximate result.

3. PRINCIPLES OF THE PROCEDURE

The EZ-SCREEN® Cup Drugs of Abuse Test is a rapid, competitive, membrane-based immunochromatographic assay. A single urine sample can be evaluated for the presence of each of the specified classes of drugs in a single device. The device consists of a control line, drug-conjugates and antibody-colloidal gold.

1. CONTROL LINE -- Each test strip has anti-mouse immunoglobulin antibody immobilized as a line on the membrane at the Control (C) location on the device window. The anti-mouse immunoglobulin antibody can bind to any of the mouse antibodies coated on the colloidal gold.

2. DRUG-CONJUGATES -- Drug from the class tested was individually conjugated to a protein that binds to the membrane. Each drug conjugate was immobilized as a line at a labeled location on the membrane strip.

3. ANTIBODY-COLLOIDAL GOLD -- Each test uses a monoclonal antibody developed to bind to its drug class. Antibody-colloidal gold solutions were prepared by absorbing each of the individual monoclonal antibodies to colloidal gold. The colloidal gold solutions were applied to the sample pad in the drugs of abuse test.

The device can be used to detect specific classes of drugs in urine because drug(s) in the urine and the drug(s) conjugated to the protein compete to bind to the antibody-colloidal gold in a highly specific reaction. When the urine flows into the sample pads of the device, the dried antibody-colloidal gold on the sample pad(s) dissolves and the urine wicks up the white strips carrying the reddish-purple antibody-colloidal gold as a solution with it.

For specifics on crossreactivity in a drug class see the Related and Reactive Compounds section for that test.

Control Line

Each test strip has an internal procedural control. A line must form at the Control "C" location on the device to indicate that the proper sample volume was used and that the reagents are migrating properly. If a Control line does not form, the test is considered invalid. A Control line forms when the antibody-colloidal gold binds to the anti-mouse immunoglobulin antibody immobilized on the membrane at the "C" location on the device.

Negative Samples

When no drug(s) is present in the urine sample, the reddish purple antibody-colloidal gold solutions migrate along the strip then binds to the appropriate drug conjugate immobilized on the membrane. The binding of the antibody-colloidal gold to the drug conjugate generates an easily visible reddish-purple line at the appropriate test location on the device. Negative results can be reported as soon as a test line and control line are visible.

Presumptive Positive Samples

When drug(s) is present in the urine sample the antibody-colloidal gold binds to the drug(s) before it migrates along the strip. However, when the antibody-colloidal gold binds to the drug(s) in the urine, the antibody colloidal gold cannot bind to the drug conjugate immobilized on the membrane. When the drug concentration is at or above the cutoff concentration, the majority of the antibody-colloidal gold is bound to the drug from the urine. Therefore, as the drug bound antibody-colloidal gold migrates along the strip(s), it is unable to bind to the appropriate drug conjugate immobilized on the membrane. Therefore no line is generated at the drug-specific location on the device for a positive sample. Read non-negative results at 5 minutes. The control line should be present for the test to be valid. The test result for Oxycodone (OX) after 5 minutes may not be consistent with the original reading. For all other tests, read results at 5 minutes or within 15 minutes of the sample application. The test result after 15 minutes may not be consistent with the original reading.

4. MATERIALS PROVIDED/STORAGE CONDITIONS

Each EZ-SCREEN® Cup Drugs of Abuse Test contains all the reagents necessary to test one urine sample simultaneously for multiple drugs. Test devices are available in Cup format as described below.

Kit Contents – Cup Test format

Each Cup Test Kit contains twenty-five (25) test system foil pouches, and one reference guide. Products with adulterant strips contain five (5) Color Comparator Charts.

Each Cup Test system foil pouch contains:

1. One (1) test cup with temperature strip attached, and test cassette(s) inside.
 1. Each test cassette has test strips with drug specific reagents.
 2. The test cup may contain a membrane strip laminated with adulterant pads for testing the presence of Oxidants and Nitrites, as well as determining approximate values of Specific Gravity and pH in human urine. (Products with adulterant strips only; the adulterant strip is not contained in every EZ-SCREEN® Cup product.)
2. One (1) lid.

Materials Required but not provided

External controls

Timer

Specimen containers, external controls, disposable gloves and urine temperature strips are available from MEDTOX Diagnostics, Inc.

Storage Conditions

The kit, in its original packaging, should be stored at 2-25°C (36-77°F) until the expiration date on the label.

5. PRECAUTIONS

1. Urine specimens and all materials coming in contact with them should be handled and disposed of as if infectious and capable of transmitting infection. Avoid contact with broken skin.
2. Avoid cross-contamination of urine samples by using a new urine specimen container for each urine sample.
3. The device should remain in its original sealed foil pouch until ready to use. If the pouch is damaged, do not use the test.
4. Do not store the test kit at temperatures above 25°C (77°F).
5. If devices have been stored refrigerated, bring to ambient temperature (18-25°C/ 64-77°F) prior to opening foil pouch.
6. Do not use tests after the expiration date printed on the package label.
7. The drug screen portion of the device is for *in vitro* diagnostic use only. The adulterant strip is for forensic/toxicology use only.

6. SAMPLE COLLECTION AND PREPARATION

Collect the urine sample in the EZ-SCREEN® Cup. The urine volume should be at or above the minimum volume fill line.

This volume of urine is more than sufficient for testing. No preservatives should be added. Tests will begin developing immediately following collection of urine. If sample needs to be confirmed it can be shipped directly to the confirmation lab in the EZ-SCREEN® Cup.

7. TEST PROCEDURE

Cup Test

1. Bring pouched cup device to room temperature before opening it.
2. Open pouch and label the device with the patient or sample identification.
3. Remove desiccant from cup.
4. Fill urine sample cup to at least the Fill Line.
5. Tighten lid onto the cup.
6. Keep cup in upright position and minimize handling before reading.
7. If adulterant strip is present, read pH, Specific Gravity, and Nitrites in vertical position as soon as color changes. Read oxidant at 60 seconds.
8. Allow the test cup to sit for 5 minutes after voiding into the cup.
9. Remove the privacy tab and read the results. Control line (C) must be present to read results.
10. If remove privacy tab before 5 minutes, negative results can be read as soon as a test line and control line (C) is visible and presumptive positive at 5 minutes after voiding into the cup.

NOTE: Read results at 5 minutes or within 15 minutes of voiding into the cup. Oxycodone should be read at 5 minutes. Test results interpreted after 15 minutes (for Oxycodone after 5 minutes) may not be consistent with the original results obtained at 5 minutes.

8. READING THE TEST RESULTS

Invalid: The absence of a reddish-purple Control (C) line at 5 minutes indicates the test is invalid. The urine sample should be retested on a new device. If the second test is also invalid, send the urine sample to a reference laboratory for additional testing.


Negative: The appearance of both a reddish-purple Control (C) line and a specific drug line (AM, BA, BZ, CO, mA, MT, OP, OX, PC, PP or TH) indicates a negative test result. The color intensities of the Control line (C) and a specific drug line may not be equal; any reddish-purple line visible at 5 minutes indicates a negative test result for that drug. Line intensity will vary from test to test.

Presumptive Positive: The appearance of both a reddish-purple Control (C) line and the absence of a line next to a specific drug name (AM, BA, BZ, CO, mA, MT, OP, OX, PC, PP or TH) at 5 minutes indicates a presumptive positive test result for that drug. Occasionally a white line (line lighter than the background of the strip) may appear next to a specific drug name. This indicates a presumptive positive test result for that drug.

Examples of Negative, Presumptive Positive and Invalid results:



INVALID
(C line is missing)



NEGATIVE



**PRESUMPTIVE
POSITIVE**
(OP Test)

There are other possible results depending on the drug or combination of drugs present in the urine sample.

9. INTERPRETATION OF TEST RESULTS

A **NEGATIVE** test result for a specific drug indicates that the sample does not contain the drug/drug metabolite above the cutoff level.

A **PRESUMPTIVE POSITIVE** test result for a specific drug indicates that the sample may contain drug/drug metabolite near or above the cutoff level. It does not indicate the level of intoxication or the specific concentration of drug in the urine sample. Presumptive Positive samples or those with abnormal adulterant strip results should be sent to a reference laboratory for more definitive testing.

10. QUALITY CONTROL

An internal procedural control is included on each test strip. A line must form at the Control (C) position in the result window to indicate that the proper sample volume was used and that the reagents are migrating properly. If a Control line does not form, the test is considered invalid. The Control line consists of immobilized anti-mouse antibody that reacts with the antibody-colloidal gold as it passes this region of the membrane. Formation of a visible line verifies the Control line antibody antigen reaction occurred. This line may be considered an internal negative procedural control. In addition, if the test has been performed correctly and the device is working properly, the background will clear such that result lines are distinct. The cleared background may be considered an internal positive procedural control. The visible Control line (C) should always be present regardless of whether drug is absent or present in the sample.

The purpose of quality control in laboratory testing is to ensure accuracy, reliability of results and to detect errors. Because the devices are self-contained, single use tests, traditional quality control programs do not apply. The Quality Control program MEDTOX recommends for these non-instrumented test devices includes a combination of the internal device controls and external controls to ensure accuracy, reliability and to detect possible errors. The on-board reactive device controls may be one aspect of the quality program utilized by a laboratory to satisfy the daily quality control requirement established by the Laboratory Director. Another aspect of a quality control program includes an external negative control containing no drug and a positive drug control challenging to the assay cutoff concentration. These controls may be used to initially test each shipment of product received by the laboratory or to verify appropriate storage conditions and long-term stability of the test reagent. To follow good laboratory practices, we recommend that the user document the receipt of each new lot number of devices, the results of external controls performed initially and periodically thereafter, and the results of the internal controls within each device.

It is the responsibility of each Laboratory Director to demonstrate and document the validity of the alternate QC procedure they choose to use in their laboratory. For additional information or forensic and workplace testing requirements, users should contact and follow the appropriate federal, state, and local guidelines. Quality control materials are available from MEDTOX and commercial sources. Contact MEDTOX for further information.

11. LIMITATIONS OF THE PROCEDURE

1. The EZ-SCREEN[®] Cup Drugs of Abuse Test is only for use with unadulterated human urine samples collected in the EZ-SCREEN[®] Cup. Urine samples which are either extremely acidic (below pH 4.0) or basic (above pH 9.0) may produce erroneous results.
2. Urine samples which are collected in another cup and then poured into an EZ-SCREEN[®] Cup may produce erroneous results.
3. Keep the EZ-SCREEN[®] Cup upright while strips are developing. Turning the EZ-SCREEN[®] Cup upside down or on its side may produce erroneous or invalid results.
4. Shaking or excessive agitation of the EZ-SCREEN[®] Cup may produce erroneous or invalid results.
5. A positive result for any drug(s) does not indicate or measure intoxication. It only indicates the presence of reacting compound(s) in the urine specimen.
6. Test results interpreted after 15 minutes (5 minutes with OX) may not be consistent with the original result obtained at 5 minutes.
7. The Drugs of Abuse Test was not evaluated in point-of-care settings.
8. There is a possibility that other substances and/or factors, e.g. technical or procedural errors, may interfere with the test and cause false results.

Adulterant Strip limitations

The purpose of the adulterant strip is to screen for abnormal conditions in human urine samples, such as dilution or the addition of drug-test interfering substances. Occasionally medications may discolor the urine, and make it difficult to read the result. When in doubt send the urine sample to a reference laboratory for additional testing.

Oxidant

Nitrites, acting as oxidizing agents in solution, will produce a blue/green color change on the Oxidant pad.

Nitrite

Abnormal results can be caused by the presence of diagnostic or therapeutic dyes in the urine. Very high concentrations of oxidant such as 80% bleach will produce a brown color change on the Nitrite pad.

12. EXPECTED VALUES

The Substance Abuse and Mental Health Services Administration (SAMHSA) recommends the following screening test cutoffs:

AM	Amphetamine	1000 ng/mL
CO	Benzoyllecgonine	300 ng/mL
mA	Methamphetamine	1000 ng/mL
OP	Morphine	2000 ng/mL
PC	Phencyclidine	25 ng/mL
TH	11-nor-9-carboxy- Δ^9 -THC	50 ng/mL

There are no SAMHSA recommended screening levels for barbiturates, benzodiazepines, MDMA, methadone, oxycodone and propoxyphene and/or their metabolites.

The Drugs of Abuse Test qualitatively detects amphetamine, barbiturates, benzodiazepines, cocaine, methadone, methamphetamine, opiates, oxycodone, phencyclidine, propoxyphene, and cannabinoids and/or their metabolites in human urine at or above their specified cutoff level (See Specificity).

Adulterant Strip

Urines that produce an abnormal result on the adulterant strip should be sent to a reference laboratory for more definitive testing to determine if the urine may be dilute, substituted, invalid and/or adulterated.

13. PERFORMANCE CHARACTERISTICS

Sensitivity

The Drugs of Abuse Test detects one or more of the following drugs at cutoff levels listed below. Cutoffs for amphetamines, cocaine metabolite, methamphetamines, opiates, phencyclidine, and cannabinoids (THC) are based on SAMHSA recommendations for screening of these drugs in human urine. There are no SAMHSA recommended screening cutoff levels for barbiturates, benzodiazepines, methadone, oxycodone, norpropoxyphene, or propoxyphene.

AM	Amphetamine	1000 ng/mL
BA	Barbiturates (Butalbital)	200 ng/mL
BZ	Benzodiazepines (Nordiazepine)	300 ng/mL
CO	Benzoyllecgonine	300 ng/mL
mA	Methamphetamine	1000 ng/mL
MT	Methadone	300 ng/mL
OP	Morphine	2000 ng/mL
OX	Oxycodone	100 ng/mL
PC	Phencyclidine	25 ng/mL
PP	Propoxyphene (Norpropoxyphene)	300 ng/mL
TH	11-nor-9-carboxy- Δ^9 -THC	50 ng/mL

Accuracy

A panel of naturally metabolized urine samples for the following drug(s) was analyzed using the MEDTOX[®] AM, CO, OP, PC and TH Tests and the Boehringer Mannheim qualitative CEDIA[®] assay or the ROCHE ABUSCREEN ONLINE[®] for each drug and the results were compared. Results are shown in the following tables.

ACCURACY COMPARED TO THE BOEHRINGER MANNHEIM QUALITATIVE CEDIA[®] or THE ROCHE ABUSCREEN ONLINE[®] II ASSAYS

CEDIA AMPHETAMINE (1000 ng/mL cutoff)

AM (1000 ng/mL cutoff)		<u>Positive</u>	<u>Negative</u>	<u>TOTAL</u>
	Positive	64	0	64
	<u>Negative</u>	<u>2</u>	<u>618</u>	<u>620</u>
	TOTAL	66	618	684

Overall agreement: >99% (682/684). Samples having discrepant results were analyzed by GC/MS. The two false negative samples contained amphetamine at 2353 and 3569 ng/mL.

CEDIA COCAINE (300 ng/mL cutoff)

CO (300 ng/mL cutoff)		<u>Positive</u>	<u>Negative</u>	<u>TOTAL</u>
	Positive	96	8	104
	<u>Negative</u>	<u>2</u>	<u>578</u>	<u>580</u>
	TOTAL	98	586	684

Overall agreement: 99% (674/684). Samples having discrepant results were analyzed by GC/MS. Of the eight false positive samples one contained 151 ng/mL while seven did not contain cocaine metabolite detectable at the GC/MS cutoff of 150 ng/mL. The two false negative samples contained cocaine metabolite at 688 and 666 ng/mL.

ROCHE ABUSCREEN ONLINE[®]-II OPIATE (2000 ng/mL cutoff)

OP (2000 ng/mL cutoff)		<u>Positive</u>	<u>Negative</u>	<u>TOTAL</u>
	Positive	68	0	68
	<u>Negative</u>	<u>0</u>	<u>89</u>	<u>89</u>
	TOTAL	68	89	157

Overall agreement: 100% (157/157).

CEDIA PHENCYCLIDINE (25 ng/mL cutoff)

		<u>Positive</u>	<u>Negative</u>	<u>TOTAL</u>
PC (25 ng/mL cutoff)	Positive	56	2	58
	<u>Negative</u>	<u>1</u>	<u>625</u>	<u>626</u>
	<u>TOTAL</u>	<u>57</u>	<u>627</u>	<u>684</u>

Overall agreement: >99% (681/684). Samples having discrepant results were analyzed by GC/MS. The two false positive samples did not contain phencyclidine detectable at the GC/MS cutoff of 25ng/mL. The one false negative sample contained phencyclidine at 28 ng/mL.

CEDIA MULTI-LEVEL THC (50 ng/mL cutoff)

		<u>Positive</u>	<u>Negative</u>	<u>TOTAL</u>
TH (50 ng/mL cutoff)	Positive	194	3	197
	<u>Negative</u>	<u>10</u>	<u>477</u>	<u>487</u>
	<u>TOTAL</u>	<u>204</u>	<u>480</u>	<u>684</u>

Overall agreement: 98% (671/684). Samples having discrepant results were analyzed by GC/MS. The three false positive samples were found to contain 16, 28, and 32 ng/mL while the ten false negative samples contained 32, 35, 41, 42, 46, 46, 49, 50, 50, and 90 ng/mL.

RELATIVE SENSITIVITY AND SPECIFICITY COMPARED TO THE BOEHRINGER MANNHEIM QUALITATIVE CEDIA® or THE ROCHE ABUSCREEN ONLINE® II ASSAYS (Amphetamines, Cocaine, Opiates, PCP and THC)

	<u>Relative Sensitivity</u>	<u>Relative Specificity</u>
AM	97% (64/66)	100% (618/618)
CO	98% (96/98)	99% (578/586)
OP	100% (68/68)	100% (89/89)
PC	98% (56/57)	>99% (625/627)
TH	95% (94/204)	99% (477/480)

ACCURACY COMPARED to GC/MS

		<u>MEDTOX® Tests</u>	<u>GC/MS</u>	<u>Values for discrepant Samples (ng/mL)</u>
AM	Positive	48	50	
	Negative	52	50	2353 and 3569
CO	Positive	49	50	
	Negative	51	50	666
OP	Positive	47	47	
	Negative	0	0	No Discrepant
PC	Positive	49	50	
	Negative	51	50	28
TH	Positive	48	50	
	Negative	52	50	35 and 46

Precision (Amphetamines, Cocaine, Opiates, Phencyclidine, and Cannabinoids)

Performance around the specific cutoff for each drug was measured by testing standard drug solutions diluted in drug-free urine in replicates of 20 each on 3 different days by 3 operators. Twenty replicates of drug-free urine were also tested on each day. At 25% above the cutoff, the precision of each assay was as follows: AM=100%, CO=100%, OP=96.7%, PC=100% and TH=95%.

Reproducibility (Amphetamines, Cocaine, Phencyclidine, and Cannabinoids)

A panel of 55 naturally metabolized human urine samples was prepared. All samples in the panel had been screened for the presence or absence of amphetamines, cocaine, phencyclidine, and cannabinoids. In addition, each of the 55 samples had also been quantitated by GC/MS conducted at SAMHSA cutoffs for positive samples or at limit of quantitation for negative samples to determine the concentration of a specific drug. Five of the 55 samples were drug-free negatives and 50 of the samples were positive for one or more of the four drugs above. The concentration of primary metabolite in the positive samples was between 1056 and 4622 ng/mL for AM; 487 and 1342 ng/mL for CO; 32 and 109 ng/mL for PC and 66 and 198 ng/mL for TH. The panel was used to evaluate the lot-to-lot and lab-to-lab reproducibility.

Lot-to-Lot Reproducibility (Amphetamines, Cocaine, Phencyclidine, and Cannabinoids)

Three aliquots of each of the 55 samples were prepared and each of the three sets of aliquots was coded and used to evaluate the performance of one of three lots of drug tests for the four drugs above. There was one incorrect result (a false negative on an amphetamine low positive sample) on the 825 tests for a reproducibility of >99%.

Lab-to-Lab Reproducibility (Amphetamines, Cocaine, Phencyclidine, and Cannabinoids)

Three aliquots of each of the 55 samples were prepared and each of the three sets of aliquots was tested by one of three study participants using one lot of drug test for the four drugs above. There was >99% agreement between the three participants. Overall, there was one incorrect result (for PC, one false negative on a low positive sample) on the 825 tests.

Reproducibility (Opiates 2000)

A panel of 25 naturally metabolized human urine samples was prepared. All samples in the panel had been screened for the presence or absence of opiates. In addition, each of the positive samples had also been quantitated by GC/MS conducted at SAMHSA cutoff for positive samples to determine the concentration of morphine and codeine. The concentration of morphine and/or codeine in the positive samples was between 2000 and 6000 ng/mL. The panel was used to evaluate Opiates 2000 for lot-to-lot and lab-to-lab reproducibility. There were no incorrect results on the 75 tests (25 samples x 3 lots) for a lot-to-lot reproducibility of 100%. There were no incorrect results on the 75 tests (25 samples x 3 study participants) for a lab-to-lab reproducibility of 100%.

Accuracy (Propoxyphene)

One-hundred forty one (141) clinical samples were evaluated by the Roche Abuscreen OnLine Propoxyphene assay, using a 300 ng/mL cut off. Sixty (60) samples were found to be negative and eighty-one (81) samples were found to be positive by the Roche method. Three aliquots of each sample were prepared, and assayed by three operators in a masked manner. There was no significant difference in the results obtained by the three operators, therefore the results of all three operators are included in the table. Results of this comparison are as follows:

	<u>OnLine Positive</u>	<u>OnLine Negative</u>
PP (300 ng/mL cutoff)	238 5*	0 180

* GC/MS results are 390, 441, 499, 536 and 679 ng/mL

In addition to the 141 clinical samples, eight additional clinical samples containing only norpropoxyphene were diluted with drug-free urine in order to obtain an adequate number of samples that had concentrations of drug that were challenging to the cutoff. These eight diluted samples, and the 141 clinical samples described above were analyzed by GC/MS for propoxyphene and norpropoxyphene. The level of quantitation of the GC/MS was 30 ng/mL. Only ten of the samples contained propoxyphene, and each of these samples had norpropoxyphene levels greater than 1,647 ng/mL. As in the study above, three aliquots of the 149 samples were prepared, coded, and assayed by three operators in a masked manner. There was no significant difference in the results obtained by the three operators, therefore the results of all three operators are included in the comparison table.

<u>GC/MS Range (ng/mL)</u>	<u>None detected</u>	<u>150-265</u>	<u>339-450</u>	<u>>472</u>
Number of samples	60	8 (Diluted samples)	7	74
Positive	0	12	19	219
Negative	180	12	2	3

Sensitivity/Precision/Distribution of Random Error (Propoxyphene)

Performance around the specific cut-off of 300 ng/mL for norpropoxyphene was evaluated by testing standard drug solutions diluted in drug-free urine in triplicate on 5 different days by 3 operators. Drug-free urine was also tested on each day. There was no significant difference in the results of the three operators so the results were combined and are shown in the following table.

<u>Conc. (ng/mL)</u>	<u>Number Tested</u>	<u>Norpropoxyphene Cut-off = 300 ng/mL</u>		<u>% Agreement</u>
		<u>Positive</u>	<u>Negative</u>	
0	45	0	45	100
30	45	0	45	100
75	45	1	44	98
150	45	9	36	80
225	45	16	29	64
300	45	37	8	82
375	45	42	3	93
450	45	44	1	98
600	45	45	0	100

Accuracy (Methamphetamine)

A panel of naturally metabolized urine samples was analyzed using the MEDTOX[®] mA Test and the GC/MS assay for methamphetamine. The results obtained in the procedures are shown in the following tables.

GC/MS Methamphetamine (limit of quantitation 50 ng/mL)

		<u>Positive</u>	<u>Negative</u>	<u>TOTAL</u>
mA-Methamphetamine (1000 ng/mL cut-off)	Positive	56	0	56
	<u>Negative</u>	<u>2</u>	<u>56</u>	<u>58</u>
	TOTAL	58	56	114

Overall agreement: >98% (112/114). Samples having discrepant results were analyzed by GC/MS. The false negative samples contained methamphetamine at 1056 ng/mL and at 1136 ng/mL.

Percent Agreement of mA Compared to GC/MS

	<u>POSITIVE</u>	<u>NEGATIVE</u>
mA-Methamphetamine	97% (56/58)	100% (56/56)

Sensitivity/Precision (Methamphetamine)

Performance for the MEDTOX[®] mA Test for methamphetamine was evaluated by testing standard drug solutions diluted in drug-free urine in duplicates of 8 drug concentrations on 5 different days by 3 operators. Drug-free urine was also tested on each day. The complete results for both drugs are shown in the tables below.

Methamphetamine Cut-off = 1000 ng/mL

<u>Conc. (ng/mL)</u>	<u>No. Tested</u>	<u>(+)</u>	<u>(-)</u>	<u>% Agreement</u>
0	30	0	30	100
100	30	0	30	100
250	30	0	30	100
500	30	26	4	87
750	30	27	3	90
1000	30	28	2	93
1250	30	29	1	97
1500	30	30	0	100
2000	30	30	0	100

Reproducibility (Methamphetamine)

A panel of 18 spiked human urine samples, comprised of drug-free and drug standard samples, was prepared. The panel was examined by 3 operators, once a day for 5 days. The concentration of methamphetamine had been quantitated by GC/MS in each of the 18 samples. There was 100% agreement between the three operators over the 5 day period at 0 ng/mL, 1500 ng/mL (cut-off + 50%) and 2000 ng/mL (cut-off + 100%) for methamphetamine.

Accuracy (Barbiturates, Benzodiazepines and Methadone)

The accuracy was evaluated by assaying a coded panel of clinical urine samples containing varying concentrations of drugs and comparing the results to validated methods. Validated GC/MS assays measured barbiturates, benzodiazepines, and methadone levels. Results are shown in the following tables.

**ACCURACY COMPARED TO GC/MS
(Barbiturates, Benzodiazepines, and Methadone)**

DRUG CLASS	Concentration Range (ng/mL)	Number of Samples	MEDTOX® Results
Barbiturates	201 – 27776	36	36/36 Positive
	155, 155, 156, 158, 161	5	5/5 Negative
Butalbital	240 - 3814	27	27/27 Positive
	109, 151, 194	3	3/3 Positive
Pentobarbital	264	1	1/1 Positive
Benzodiazepines	303 – 30813	57	57/57 Positive
	234, 236, 238, 250, 283	5	5/5 Negative
Methadone	306 - 70560	57	57/57 Positive
	224, 226, 227, 230, 232	5	5/5 Negative

Additionally, the accuracy was evaluated in comparison to the Roche Diagnostics Systems, Inc., ABUSCREEN ONLINE® assays for barbiturates, benzodiazepines and methadone. A panel of clinical urine samples was analyzed and the results obtained in the procedures were compared. Results are shown in the following tables.

**ACCURACY COMPARED TO THE ROCHE ABUSCREEN ONLINE® II
(Barbiturates, Benzodiazepines, and Methadone)**

ABUSCREEN ONLINE® II Barbiturates Result (Secobarbital)
(300 ng/mL cutoff)

		<u>Positive</u>	<u>Negative</u>	<u>TOTAL</u>
BA (200 ng/mL cutoff)	Positive	62	0	62
Butalbital Test	Negative	0	45	46
	TOTAL	62	45	107

Overall agreement: 100% (107/107).

ABUSCREEN ONLINE® II Benzodiazepines Result
(300 ng/mL cutoff)

		<u>Positive</u>	<u>Negative</u>	<u>TOTAL</u>
BZ (300 ng/mL cutoff)	Positive	57	0	57
Nordiazepam Test	Negative	0	45	45
	TOTAL	57	45	102

Overall agreement: 100% (102/102).

ABUSCREEN ONLINE® II Methadone Result
(300 ng/mL cutoff)

		<u>Positive</u>	<u>Negative</u>	<u>TOTAL</u>
MT (300 ng/mL cutoff)	Positive	55	0	55
Methadone Test	Negative	0	45	45
	TOTAL	55	45	100

Overall agreement: 100% (100/100).

**PERCENT AGREEMENT COMPARED TO ROCHE ABUSCREEN ONLINE ASSAYS
(Barbiturates, Benzodiazepines, and Methadone)**

	<u>POSITIVE</u>	<u>NEGATIVE</u>
Barbiturates	100% (62/62)	100% (45/45)
Benzodiazepines	100% (57/57)	100% (45/45)
Methadone	100% (55/55)	100% (45/45)

Sensitivity / Precision / Distribution of Random Error (Barbiturates, Benzodiazepines, and Methadone)

Performance around the specific cutoff for each drug was evaluated by testing standard drug solutions diluted in drug-free urine in triplicate on 5 different days by 3 operators. Drug-free urine was also tested on each day. Operator-to-operator agreement was excellent; therefore, the data were combined and summarized in the following tables.

Conc. (ng/mL)	Number Tested	Barbiturates (Butalbital) Cutoff = 200 ng/mL		% Agreement
		Positive	Negative	
Negative	45	0	45	100
50	45	0	45	100
100	45	0	45	100
150	45	12	33	73
200	45	43	2	96
250	45	45	0	100
300	45	45	0	100

Benzodiazepines (Nordiazepam) Cutoff = 300 ng/mL

Conc. (ng/mL)	Number Tested	Positive	Negative	% Agreement
Negative	45	0	45	100
30	45	0	45	100
75	45	6	39	87
150	45	27	18	60
225	45	41	4	91
300	45	42	3	93
375	45	43	2	96
450	45	45	0	100
600	45	45	0	100

Methadone (Methadone) Cutoff = 300 ng/mL

Conc. (ng/mL)	Number Tested	Positive	Negative	% Agreement
Negative	45	0	45	100
30	45	3	42	93
75	45	28	17	62
150	45	35	10	78
225	45	43	2	96
300	45	45	0	100
375	45	45	0	100
450	45	43	2	96
600	45	44	1	98

Accuracy (Oxycodone)

The accuracy was evaluated by assaying a panel of blind coded clinical urine samples containing varying concentrations of drugs and comparing to GC/MS results. The samples were obtained from MEDTOX Laboratories. Samples that screened negative by the predicate device were not confirmed by GC/MS. Positive samples were confirmed by GC/MS. The GC/MS determination included Oxycodone and oxymorphone and a weighted concentration using 100% cross-reactivity for Oxycodone and a 50% cross-reactivity for oxymorphone was calculated. Clinical urine samples containing Oxycodone and oxymorphone at higher concentrations were diluted with negative urine to obtain the desired number of samples with concentrations below and above the cutoff. The testing was performed by nine MEDTOX personnel at one site.

MEDTOX® OXYCODONE Results vs. stratified GC/MS Values

MEDTOX® OXYCODONE Results	Negative by Immunoassay (Predicate Device)	Concentration up to 50% below the cutoff	Near Cutoff Negative (Between 50% below the cutoff and the cutoff concentration)	Near Cutoff Positive (Between the cutoff and 50% above the cutoff concentration)	High Positive (Greater than 50% above the cutoff concentration)
Positive	0	2	2	6	38
Negative	103	5	4	1	0

GC/MS values used to categorize samples in this table are determined by adding together the concentration of Oxycodone plus 50% of the concentration of oxymorphone, based on the MEDTOX® OXYCODONE cross-reactivity studies.

% Agreement among positives is 98%. % Agreement among negatives is 97%.

Sensitivity/Precision (Oxycodone)

Performance around the specific cutoff for Oxycodone was evaluated by testing standard drug solutions diluted in drug-free urine in triplicate on 6 different intervals by 3 in-house operators. Drug free urine was also tested on each interval. The results were interpreted at five minutes and are summarized below:

MEDTOX® OXYCODONE Precision Study Results

Concentration of sample (ng/mL)	Number of determinations	Results #Neg / #Pos
0	54	54 / 0
25	54	54 / 0
50	54	50 / 4
75	54	14 / 40
100	54	4 / 50
125	54	1 / 53
150	54	0 / 54

Non Crossreactive Endogenous Compounds

Listed compounds were dissolved in appropriate solvents and then added to drug-free urine for evaluation with all tests. Most of the compounds were evaluated for reactivity at 100 µg/mL (albumin was evaluated at 20 mg/mL and bilirubin was evaluated at 200 µg/mL). Samples were evaluated in triplicate by in-house operators. None of these compounds showed cross-reactivity at the listed concentrations.

Acetaldehyde	Creatinine	Hemoglobin, Human
Acetone	Epinephrine	Sodium Chloride
Albumin, Human	β-Estradiol	Tetrahydrocortisone
Bilirubin	Estriol	d,1-Thyroxine
Cholesterol	Glucose Std. Solution	Uric Acid

Unrelated Compounds, Prescription and Over-the-Counter Medications

The following compounds were tested for reactivity. Listed compounds were dissolved in appropriate solvents and then added to drug-free urine for testing. Samples were evaluated in triplicate by in-house operators. Unless otherwise noted by a drug name abbreviation such as "AM" or "BA" etc., all of the listed compounds were negative in each of the tests at 100 µg/mL or the highest level tested. If a drug name is followed by an abbreviation such as "AM" or "BA" etc., check the "Related Compounds and Cross Reactants" listing for the drug in question under the appropriate heading (AM, BA, etc.) to find its level of cross-reactivity to that test.

Acetaminophen	Acetylsalicylic Acid
Alphenal- BA	Alprazolam- BZ
Alprazolam, 1-Hydroxy- BZ	7-Aminoclonazepam
7-Aminoflunitrazepam	1-Aminopyrine (4-(dimethylamino) antipyrine)
Amitriptyline	Amoxapine
Amoxicillin	1-Amphetamine- AM
Ampicillin	l-Ascorbic Acid
Aspartame	Atomoxetine
Atropine Sulfate	Barbituric Acid
Benzilic Acid	Benzocaine (ethyl-4-aminobenzoate)
Benzoyllecgonine- CO	Benztropine
Brompheniramine	Bupropion
Butabarbital- BA	Caffeine
Cannabidiol	Captopril
Carbamazepine	Carisoprodol (Meprobamate)
Cephalexin	Chloramphenicol
Chlordiazepoxide	Chlorothiazide
Chlorpheniramine	Chlorprothixene
Clobazam- BZ	Clonazepam- BZ
Clonidine	Clozapine
Cocaine- CO	Cortisone
Cotinine	Cyclopentobarbital- BA
Deoxycorticosterone	Desipramine
Desmethylchlordiazepoxide- BZ	Desmethylvenlafaxine
Dexamethasone	Diacetylmorphine- OP
Diazepam- BZ	Diethylpropion
Diflunisal	Dihydrocodeine- OP, OX
Dimenhydrinate (Dramamine)	Diphenhydramine
Diphenylhydantoin (Phenytoin)- BA	Dopamine
Doxepin	Ecgonine
EDDP-(Primary metabolite of methadone)	EMDP-(Secondary metabolite of methadone)
Ephedrine- mA	Erythromycin
Estrone	Ethylmorphine- OP, OX
Fenfluramine- mA	Fentanyl (Synthetic opiate)
Flunitrazepam- BZ	Flurazepam
Furosemide	Genistic Acid (2,5-Dihydroxybenzoic acid)
Glutethimide	Haloperidol
Hexobarbital	Hyalalazine
Hydrochlorothiazide	Hydrocortisone
Hydromorphone- OP, OX	Hydroxyhippuric Acid
l-11-Hydroxy- Δ^9 -THC	4-Hydroxyphenacylidine- PC
3-Hydroxytyramine	Ibuprofen
Imipramine	(R)-Isoproterenol
Isoxsuprine- CO	Ketoprofen
Labetalol	Lidocaine
Lithium carbonate	Lorazepam- BZ
Lorazepam glucuronide- BZ	Lysergic Acid
Lysergic Acid Diethylamide (LSD)	MDA- AM
MDE (MDEA)- mA	Melanin
Meperidine	Mepivacaine
Mesoridazine	d-Methamphetamine- mA
l-Methamphetamine- mA	Methcathinone
Methocarbamol	Methylphenidate
Methylprylon	Midazolam- BZ
Mirtazapine	Morphine- OP, OX
Morphine 3- β -D-Glucuronide- OP	Nalidixic Acid
Naltrexone- OX	Naloxone- OX
Naproxen	Nicotine
Nifedipine	Nitrofurantoin
Norclomipramine	Nordiazepam- BZ
Nordoxepin	Norlysergic Acid
Normeperidine	l-Norpseudoephedrine
Nortriptyline	Nylidrin
Octopamine	Olanzapine
Omeprazole	Oxalic Acid
Oxaprosin	Oxazepam glucuronide- BZ
Oxolinic Acid	Oxymetazoline
Oxymorphone- OX	Penicillin G
Pentazocine	Perphenazine
Phenacetin (Acetophenetidin)	Phendimetrazine
Phenelzine	Pheniramine
Phenmetrazine	Phenothiazine
Phentermine- AM	Phenylbutazone
Phenylephrine	Piroxicam
Prazosin	Prednisone
Procaine	Prochlorperazine
Promazine	Propoxyphene- PP
Propranolol	d-Pseudoephedrine
Pyrilamine	Quinidine
Ranitidine	Rifampin
Salicylic Acid	Selegiline (Deprenyl)
Serotonin (5-Hydroxytryptamine)	Sildenafil (Viagra)
Sulfamethazine	Talbutal- BA
Temazepam- BZ	Tetracycline
Δ^9 -Tetrahydrocannabinol	Tetrahydrozoline
Thebaine- OP	Thiamine
Thiopental	Thiothixene

Tolbutamide
Triamterene
Trifluoperazine
Tripelethamine
Tyramine
Venlafaxine

Tolmetin (Tolectin)
Triazolam-**BZ**
Trimethoprim
Tryptamine
Tyrosine
Verapamil

Trazodone
Triazolam, 1-hydroxy
Trimipramine
Tryptophan
Valproic Acid
Zomepirac

Related Compounds and Cross Reactants

The following metabolites and compounds were tested. Reference standards for the various metabolites and compounds were prepared in negative urine samples. None of the compounds reacted with the remaining tests in the panel. Results are expressed as the minimum concentration required to produce a positive result in the indicated assay. Compounds that reacted with the test are listed first, and related compounds that did not react with the highest concentration tested are listed second as Negative at 100,000 ng/mL (or highest level tested).

Amphetamine-(AM) (d-Amphetamine) 1000 ng/mL

l-Amphetamine
MDA
Phentermine

Ephedrine
MDMA
MDE (MDEA)
l-Methamphetamine
d-Methamphetamine
Phenethylamine
Tyramine

Result

Positive at 100,000 ng/mL
Positive at 400 ng/mL
Positive at 10,000 ng/mL

Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL

Barbiturate-(BA) (Butalbital) 200 ng/mL

Allobarbitol
Alphenal
Amobarbital
Barbital
Butabarbital
Cyclopentobarbital
p-Hydroxyphenobarbital
Pentobarbital
Phenobarbital
Phenytoin (Diphenylhydantoin)
Secobarbital
Talbutal

Amino glutethimide
Barbituric Acid
1,3 Dimethylbarbituric Acid
Glutethimide
Hexobarbital
Mephobarbital

Result

Positive at 500 ng/mL
Positive at 100 ng/mL
Positive at 2500 ng/mL
Positive at 2500 ng/mL
Positive at 750 ng/mL
Positive at 250 ng/mL
Positive at 500 ng/mL
Positive at 500 ng/mL
Positive at 800 ng/mL
Positive at 2500 ng/mL
Positive at 75 ng/mL
Positive at 50 ng/mL

Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL

Benzodiazepine-(BZ) (Nordiazepam) 300ng/mL

Alprazolam
Alprazolam, 1-OH
Clobazam
Clonazepam
Clorazepate
Desalkylflurazepam
Desmethylchlordiazepoxide
Desmethylflunitrazepam
Diazepam
Flunitrazepam
Lorazepam
Lorazepam glucuronide
Midazolam
Nitrazepam
Oxazepam
Oxazepam glucuronide
Temazepam
Temazepam glucuronide
Triazolam

7-Aminoclonazepam
7-Aminoflunitrazepam
Chlordiazepoxide
Flurazepam
Triazolam, 1-OH

Result

Positive at 250 ng/mL
Positive at 25,000 ng/mL
Positive at 50 ng/mL
Positive at 250 ng/mL
Positive at 250 ng/mL
Positive at 250 ng/mL
Positive at 500 ng/mL
Positive at 75 ng/mL
Positive at 50 ng/mL
Positive at 75 ng/mL
Positive at 2,500 ng/mL
Positive at 1,000 ng/mL
Positive at 5,000 ng/mL
Positive at 50 ng/mL
Positive at 500 ng/mL
Positive at 2,500 ng/mL
Positive at 50 ng/mL
Positive at 750 ng/mL
Positive at 750 ng/mL

Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 10,000 ng/mL

Cocaine-(CO) (Benzoylcegonine) 300 ng/mL

Cocaine
Isoxsuprine

Ecgonine
Ecgonine Methyl Ester

Result

Positive at 800 ng/mL
Positive at 6,000 ng/mL

Negative at 100,000 ng/mL
Negative at 100,000 ng/mL

Methamphetamine-(MA) (d-Methamphetamine) 1000 ng/mL

Ephedrine
Fenfluramine
MDE (MDEA)
MDMA
l-Methamphetamine
Phenethylamine

d-Amphetamine
l-Amphetamine
MDA
Phentermine
Pseudoephedrine
Tyramine

Result

Positive at 2,500 ng/mL
Positive at 25,000 ng/mL
Positive at 5,000 ng/mL
Positive at 1,500 ng/mL
Positive at 7,500 ng/mL
Positive at 2,500 ng/mL

Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL

Methadone-(MT) (Methadone) 300 ng/mL

Primary metabolite (EDDP)
Secondary metabolite (EMDP)

Result

Negative at 100,000 ng/mL
Negative at 100,000 ng/mL

Opiates(2000)-(OP) (Morphine) 2000 ng/mL

Codeine
Diacetylmorphine
Dihydrocodeine
Ethylmorphine
Hydrocodone
Hydromorphone
Levorphanol
6-Monoacetyl Morphine
Morphine 3-β-D-Glucuronide
Morphine 6-β-D-Glucuronide
Norcodeine
Thebaine

Result

Positive at 800 ng/mL
Positive at 2,000 ng/mL
Positive at 3,000 ng/mL
Positive at 400 ng/mL
Positive at 2,000 ng/mL
Positive at 3,000 ng/mL
Positive at 12,500 ng/mL
Positive at 3,000 ng/mL
Positive at 3,000 ng/mL
Positive at 25,000 ng/mL
Positive at 25,000 ng/mL
Positive at 50,000 ng/mL

Apomorphine
Nalorphine
Naloxone
Naltrexone
Oxycodone
Oxymorphone

Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL

Oxycodone-(OX) (Oxycodone) 100 ng/mL

Codeine
Dihydrocodeine
Ethylmorphine
Hydrocodone
Hydromorphone
Morphine
Naloxone
Naltrexone
Norcodeine
Oxymorphone

Result

Positive at 2,500 ng/mL
Positive at 2,500 ng/mL
Positive at 2,500 ng/mL
Positive at 10,000 ng/mL
Positive at 10,000 ng/mL
Positive at 5,000 ng/mL
Positive at 10,000 ng/mL
Positive at 25,000 ng/mL
Positive at 50,000 ng/mL
Positive at 200 ng/mL

Apomorphine
Diacetylmorphine
Levorphanol
6-Monoacetylmorphine
Morphine 3-β-D-Glucuronide
Morphine 6-β-D-Glucuronide
Nalorphine
Thebaine

Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 50,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 10,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL

Propoxyphene-(PP) (Norpropoxyphene) 300 ng/mL

Propoxyphene

Result

Positive at 50 ng/mL

Phencyclidine-(PC) (Phencyclidine) 25 ng/mL

4-Hydroxyphencyclidine

Result

Positive at 5,000 ng/mL

Cannabinoids-(TH) (11-nor-9-carboxy-Δ⁹-THC) 50 ng/mL

Cannabidiol
Cannabinol
1-11 Hydroxy-Δ⁹-THC
Δ⁸-Tetrahydrocannabinol
Δ⁹-Tetrahydrocannabinol

Result

Negative at 100,000 ng/mL
Negative at 100,000 ng/mL
Negative at 50,000 ng/mL
Negative at 100,000 ng/mL
Negative at 100,000 ng/mL

Interference-Oxycodone Only

pH and Specific Gravity:

The MEDTOX[®] OXYCODONE test was assayed with six negative clinical samples with pH values of 4.0, 5.0, 6.0, 7.0, 8.0 and 9.0 ± 0.1. Each sample was assayed in triplicate. The pH samples were fortified with Oxycodone to the concentrations of 25 ng/mL and 150 ng/mL. All the pH levels gave negative results when fortified to 25 ng/mL, and all pH levels gave positive results when fortified to 150 ng/mL.

The MEDTOX[®] OXYCODONE test was assayed with eight samples with specific gravity values of 1.003, 1.005, 1.010, 1.015, 1.020, 1.025, 1.030 and 1.035 ± 0.001. Each sample was assayed in triplicate. The specific gravity samples were fortified with Oxycodone to the concentrations of 25 ng/mL and 150 ng/mL. All the specific gravity levels gave negative results when fortified to 25 ng/mL, and all specific gravity levels gave positive results when fortified to 150 ng/mL.

Common Drugs:

Following the study of M.L. Smith, et. al.⁶ drug free urine samples were spiked with Oxycodone to the concentrations of 25 ng/mL and 150 ng/mL. 100 µg/mL of the common drugs were then added to the preparation and assayed by the MEDTOX[®] OXYCODONE test. Samples were evaluated in triplicate by in-house operators. None of the common drugs listed in the following table affected the expected results.

Acetylsalicylic Acid	Chlorpheniramine	Ibuprofen
Acetaminophen	Cocaine	Morphine-OX
Brompheniramine maleate	Dextromethorphan	Phenobarbital
Caffeine	Diphenylhydantoin	d-Pseudoephedrine
Carbamazepine	Doxylamine	Salicylic Acid

Interference Propoxyphene/Methamphetamine Only

Common Drugs:

Following the study of M.L Smith, et. al.⁶ the following drugs were tested to determine the degree of interference they may have on the test. Commercial negative urine was spiked with 100 µg/mL of each of these drugs and with either 75 ng/mL or 600 ng/mL of norpropoxyphene or methamphetamine. Each spiked sample was tested in triplicate on the test. None of these drugs affected the expected negative or positive results with either the 75 ng/mL or 600 ng/mL fortified samples. The drugs are listed below.

Acetylsalicylic Acid	Chlorpheniramine	Ibuprofen
Acetaminophen	Cocaine	Morphine-
Brompheniramine maleate	Dextromethorphan	Phenobarbital
Caffeine	Diphenylhydantoin	d-Pseudoephedrine
Carbamazepine	Doxylamine	Salicylic Acid

14. BIBLIOGRAPHY

1. Blum, K. Handbook of Abusable Drugs. Gardener Press, Inc. New York, New York, 1984. pp. 305-349.
2. DeCresce, R.P., Lifshitz, M.S., Mazura, A.C. and Tilson, J.E. Drug Testing in the Workplace. ASCP Press. American Society of Clinical Pathologists. Chicago, Illinois. 1989. pp. 105-109.
3. Baselt, R.C. Disposition of Toxic Drugs and Chemicals in Man. Eighth Edition. Biomedical Publications. Foster City, California, 2008.
4. White, R.M. and Black, M.L. Pain Management Testing Reference. AACCPress. Washington, DC. 2007.
5. Cary, P.L. The Marijuana Detection Window: Determining the Length of Time Cannabinoids will Remain Detectable in Urine Following Smoking: A Critical Review of Relevant Research and Cannabinoid Detection Guidance for Drug Courts, Drug Court Review, Volume V:1. 2005, pp. 23 – 58.
6. Smith, M.L., Shimomura, E.T., Summers, J., Paul, B.D., Nichols, D., Shippee, R., Jenkins, A.J., Darwin, W.D., and Cone, E.J. Detection Times and Analytical Performance of Commercial Urine Opiate Immunoassays Following Heroin Administration, Journal of Analytical Toxicology. Volume 24:7. October 2000, pages 522-529.

15. LIMITED EXPRESS WARRANTIES

The manufacturer makes no express warranty other than the diagnostic test kit will measure certain drugs and/or drug metabolites when used in accordance with the manufacturer's printed instructions. The use of the kit for any other purpose is outside the intended use of this product. The manufacturer gives no express warranty as to what the legal or clinical significance is of the levels of drug(s)/drug metabolites detected by the MEDTOX test. The manufacturer disclaims any and all implied warranties of merchantability, fitness for use or implied utility for any other purposes. Any and all damages for failure of the kit to perform to its instructions are limited to the replacement value of the kit.

Covered by one or more patents.

U.S. Patent Nos. 6,566,051, 6,376,251, 6,653,139, 7,458,942

This product does not contain controlled substances.

This product does not contain hazardous or toxic chemicals as defined by the OSHA Hazard Communication Rule [29 CFR 1910.1200(g)].

MEDTOX Diagnostics Inc.
1238 Anthony Road
Burlington, NC 27215

To place an order or for technical services call 1-800-832-3244.

P/N 102287
Rev. 8/11
Printed in USA

© 2011 MEDTOX Diagnostics, Inc. All rights reserved.